

Oxygen Supply Times Chart

The chart below reflects the number of hours an oxygen-filled cylinder will last when used with a PulseDose® Conserving Device or a Continuous Flow Regulator. Note: All ambulatory ranges are calculated assuming a breath rate of 20 breaths per minute in PulseDose (PD) mode.

USE TIMES SHOWN IN HOURS									
Delivered Volume cc's:	16.5	24.75	33	41.25	49.5	66	82.5	99	
Flow Rate:	1	1.5	2	2.5	3	4	5	6	Mode
Walkabout Mini M-6 Cylinder 164 Gaseous Liters	2.7	1.8	1.4	1.1	.9	.7	.6	.4	CF
	8.3	5.5	4.1	3.3	2.8	2.1	1.7	1.4	PD
Hideaway ML-6 Cylinder 170 Gaseous Liters	2.8	1.9	1.4	1.1	.9	.7	.6	.5	CF
	8.6	5.7	4.3	3.4	2.9	2.1	1.7	1.4	PD
Walkabout 1 "C" Cylinder 240 Gaseous Liters	4.0	2.7	2.0	1.6	1.3	1.0	.8	.7	CF
	12.1	8.1	6.1	4.9	4.0	3.0	2.4	2.0	PD
Walkabout 2 "D" Cylinder 415 Gaseous Liters	6.9	4.6	3.5	2.8	2.3	1.7	1.4	1.2	CF
	21.0	14.0	10.5	8.4	7.0	5.2	4.2	3.5	PD
Walkabout 3 "E" Cylinder 682 Gaseous Liters	11.4	7.6	5.7	4.6	3.8	2.8	2.3	1.9	CF
	34.4	23.0	17.2	13.8	11.5	8.6	6.9	5.8	PD

CF=Continuous Flow

PD=PulseDose

